



STANDARDS:
ALL TIG WELDS CLEANED
ALL EDGES SMOOTHED
CORNERS TIG WELDED
AND GROUND SMOOTH
BEND RADIUS IS .125"

PLEASE RETURN A SIGNED COPY TO G2. EACH DRAWING IS TO BE SIGNED AND DATED INDICATING CUSTOMERS APPROVAL. G2 MUST OBTAIN APPROVAL ON DRAWINGS PRIOR TO THE FABRICATION OF ANY PRODUCTS. ALL AMENDMENTS MADE AFTER THE APPROVAL OF THESE DRAWINGS WILL BE AT THE EXPENSE OF THE CUSTOMER.

Approval: _____

Date: _____



PROPRIETARY
THIS DOCUMENT AND THE DATA DISCLOSED HEREIN OR HEREWITH IS NOT TO BE REPRODUCED, USED OR OTHERWISE DISCLOSED IN WHOLE OR IN PART TO ANYONE WITHOUT THE WRITTEN AUTHORIZATION FROM G2 AUTOMATED TECHNOLOGIES, LLC

A4

Approvals		Date	GAT-CAB-18.754880-2-4-SS	
Drawn By J. Lamb		9/28/11		
Approved By N/A				
TITLE SS Surgical Cabinet Double Sliding Upper Doors Double Hinged Lower Doors (3) Removable Shelf		Material: SS 304 #3 Cabinet Body and Door Mcmaster-Carr # 1726A8 SS Pull Handle Southco # EH-6A-5V5-50 Hinge Southco # J-EH-6A-10V8 SS Ball Catch		
SIZE N/A	CAGE CODE	DWG NO 18.754880-Overall		REV
SCALE Not To Scale	Qty.: 1		SHEET	